


























## Symbols Glossary








Symbol	Standard Reference	Standard Title	Symbol Title	Explanatory Text
	ASTM F2503	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.	Magnetic Resonance (MR) Unsafe	An item which poses unacceptable risks to the patient, medical staff or other persons within the MR environment.
	ASTM F2503	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.	MR Conditional	An item with demonstrated safety in the MR environment within defined conditions including conditions for the static magnetic field, the time-varying gradient magnetic fields and the radiofrequency fields.
	ASTM F2503	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.	Supplementary marking for MR Conditional Items	Additional information that, in association with an "MR Conditional" marking, states the conditions in which an item can be used safely within the MR environment
	ASTM F2503	Standard Practice for Marking Medical Devices and Other Items for Safety in the Magnetic Resonance Environment.	MR Safe	An item that poses no known hazards resulting from exposure to any MR environment. MR Safe items are composed of materials that are electrically non-conductive, non-metallic, and non-magnetic.

	IEC 60601-1, Table D.2, Symbol 10	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.	Follow instructions for use	Refer to instruction manual/booklet.
	EN 980, Clause 5.18	Symbols for use in the labelling of medical devices.	Consult instructions for use	Indicates the need for the user to consult the instructions for use.
	ISO 15223-1, Clause 5.4.3	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		
	IEC 60601-1, Table D.1, Symbol 11	Medical electrical equipment — Part 1: General requirements for basic safety and essential performance.		
	ISO 7000-1641	Graphical symbols for use on equipment.		
	EN 980, Clause 5.17.3	Symbols for use in the labelling of medical devices.	Storage temperature range	Indicates the temperature limits to which the medical device can be safely exposed.
	ISO 15223-1, Clause 5.3.7	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		
	ISO 7000-0632	Graphical symbols for use on equipment.		
	WEEE Directive 2012/19/EU	Waste and Electronic Equipment Directive	Recycle: Electronic Equipment	Separate collection for electrical and electronics equipment
	IEC 60601-1, Table D.1, Symbol 19	Graphical symbols for use on equipment.	To identify a type B applied part complying with IEC 60601-1.	To identify a type B applied part complying with IEC 60601-1.
	IEC 60601-1, Table D.1, Symbol 19	Graphical symbols for use on equipment.		

	EN 980, Clause 5.12	Symbols for use in the labelling of medical devices.	Manufacturer	Indicates the medical device manufacturer.
	ISO 15223-1, Clause 5.1.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		
	ISO 7000-3082	Graphical symbols for use on equipment		
	EN 980, Clause 5.13	Symbols for use in the labelling of medical devices.	Authorized European representative	Indicates the Authorized Representative in the European Community.
	ISO 15223-1, Clause 5.1.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		
	ISO/DIS 15223-1:2020(E) DRAFT Reference no. 5.7.7	Medical Devices — Symbols to be used with medical device labels, labeling, and information to be supplied — Part 1: General requirements. (Draft of New Version)	Medical device	Indicates the item is a medical device
	EN 980, Clause 5.2	Symbols for use in the labelling of medical devices.	Do not reuse	Indicates a medical device that is intended for one use, or for use on a single patient during a single procedure.
	ISO 15223-1, Clause 5.4.2	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		
	ISO 7000-1051	Graphical symbols for use on equipment		
	EN 980, Clause 5.10	Symbols for use in the labelling of medical devices.	Catalogue or model number	Indicates the manufacturer's catalogue number so that the medical device can be identified.
	ISO 15223-1, Clause 5.1.6	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		
	ISO 7000- 2493	Graphical symbols for use on equipment		
	EN 980, Clause 5.4	Symbols for use in the labelling of medical devices.	Batch Code	Indicates the manufacturer's batch

	ISO 15223-1, Clause 5.1.5	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		code so that the batch or lot can be identified
	ISO 7000- 2492	Graphical symbols for use on equipment		
	ISO 15223-1: 2021	Medical devices — Symbols to be used with medical device labels, labeling, and information to be supplied – Part 1: General requirements.	Unique device identifier	Indicates a carrier that contains unique device identifier information
	ISO 15223-1: 2021	Medical devices — Symbols to be used with medical device labels, labeling, and information to be supplied – Part 1: General requirements.	Serial Number	Indicates the manufacturer serial number so that a specific medical device can be identified
	ISO 15223-1: 2021	Medical devices — Symbols to be used with medical device labels, labeling, and information to be supplied – Part 1: General requirements.	Date of Manufacture	Indicates the date when the medical device was manufactured
	EN 980, Clause 5.21	Symbols for use in the labelling of medical devices.	Keep dry	Indicates a medical device that needs to be protected from moisture.
	ISO 15223-1, Clause 5.3.4	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.		
	ISO 7000-0626	Graphical symbols for use on equipment.		
	ISO 15223-1, Clause 5.3.1	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Fragile, handle with care	Indicates a medical device that can be broken or damaged if not handled carefully.
	ISO 7000-0621	Graphical symbols for use on equipment.		

	ISO 15223-1, Clause 5.3.8	Medical devices — Symbols to be used with medical device labels, labelling and information to be supplied.	Storage humidity range	Indicates the range of humidity to which the medical device can be safely exposed.
	ISO 7000-2620	Graphical symbols for use on equipment.		
	ISO 7000-0623	Graphical symbols for use on equipment.	This way up	To indicate correct upright position of the transport package.
	EU Medical Device Directive (MDD) 93/42/EEC  EU Medical Device Regulation (MDR) (2017/745)	The requirements for accreditation and market surveillance relating to the marketing of products; Medical Device Directive and European Medical Device Regulation.	CE marking	Signifies European technical conformity.
	ISO 15223-1, Clause 5.1.8	Medical devices — Symbols to be used with information to be supplied by the manufacturer.	Importer	Indicates the entity importing the medical device into the locale.
	ISO 7000-3725	Graphical symbols for use on equipment.		
 <p>eIFU indicator</p>	ISO 15223-1, Clause 5.4.3.	Medical devices — Symbols to be used with information to be supplied by the manufacturer.	Consult <i>instructions for use</i> or consult electronic <i>instructions for use</i>	Indicates the need for the user to consult the <i>instructions for use</i> .  An electronic instructions for use (eIFU) indicator (website address) may accompany the symbol when used to indicate an instruction to consult an eIFU.
FCC ID: OH2XXXX	47 CFR Part 15	Radio Frequency Devices (US).	Federal Communication Commission Number (FCC ID #)	Complies with United States Radio communication requirements.
IC: 3408B-XXXX	RSS-102	Radio Standards Specifications (CA).	Industry Canada Radio Communications License (IC:#)	Complies with Industry Canada Radio communication requirements.

	AS/NZS 4417.1:2012	Australian Radio Communications Requirements.	Australian Radio Communications License	Complies with Australian Radio communications requirements.
	N/A	N/A	U8 Safety Tested cTUV mark	Cleared for sale by TUV SUD.
	N/A	N/A	BAUART, Type Tested TUV Mark	Cleared for sale by TUV SUD.
	N/A	The Act on the Promotion of Effective Utilization of Resources and the Treatment of Specified Resources-recycled Products Stipulated in the Act	Li-ion Recycling	Indicates the presence of Lithium-ion Batteries
	ISO 7010-W012	Registered Safety Signs	General Warning Symbol	Warning Electrical Hazard
	ISO 7010-W012	Registered Safety Signs	Pinching Hazard Symbol/ Warning; Crushing of hands	To warn of a closing motion of mechanical parts of equipment
	N/A	N/A	Table Adapter Installation Symbol	Symbol used to help user properly install the table adapters.